



**CERTIFICATE OF ANALYSIS**

STP pHox Ultra / CCX Chloride Sensor  
Lot Number 25344038

Part Number 48557  
Exp Date 2027-03-10  
Date of Manufacture 2025-12-10  
UDI# (01)10385480485576(11)251210(17)270310(10)25344038

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	8.50 - 13.00	Pass
pHOx Ctl L4	121.3-130.3	Pass
pHOx Ctl L5	91.5-100.5	Pass
Std C mV	10.00 - 90.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink that reads "Olivia Ramona".

Approval

QC Inspector  
Title

1/6/2026  
Date