



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 500 Sample
Lot Number 25342043

Ref Number 57829
Exp Date 2027-05-18

Date of Manufacture 2025-11-18

UDI# (01)00385480578295(11)251118(17)270518(10)25342043

Table with 3 columns: Cal A L/N 25325072, Cal B L/N 25325073, Cal C L/N 25325074. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, pH, tHb g/dL with Spec, Pass/Fail, Assigned Value*, and Result columns.

Table with 3 columns: Cal D L/N 25325075, Flush L/N 25322032, Reference L/N 25325076. Rows include pH @ 37C, PO2, KCl with Spec, Pass/Fail, Assigned Value*, and Result columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/18/2025
Date