



CERTIFICATE OF ANALYSIS

Glucose pHox Plus L Membrane Kit
Lot Number 25205001

Part Number 33871
Exp Date 2026-10-24
Date of Manufacture 2025-07-24
UDI# (01)00385480338714(11)250724(17)261024(10)25205001

	<u>Spec*</u>	<u>Mean Result</u>
Slope	5.00 - 45.00	21.20
pHOx Plus L3	78- 92	83
Set A L1	≥470	512

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

*Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

1/13/2026
Date