



**CERTIFICATE OF ANALYSIS**

Allegro CRP Test Cartridge  
Lot Number 25293014

Ref Number 62050  
Exp Date 2026-10-20  
Date of Manufacture 2025-10-20  
UDI# (01)00385480620505(11)251020(17)261020(10)25293014

	Ctl Level 1		Ctl Level 2	
	<u>Spec</u>	<u>Pass/Fail</u>	<u>Spec</u>	<u>Pass/Fail</u>
CRP mg/L	10.6 - 24.6	Pass	50.0 - 96.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

01/07/2026  
Date