



**CERTIFICATE OF ANALYSIS**

Allegro Lipids Test Cartridge  
Lot Number 25351001

Ref Number 54652  
Exp Date 2027-06-17  
Date of Manufacture 2025-12-17  
UDI# (01)00385480546522(11)251217(17)270617(10)25351001

	Ctl Level 1				Ctl Level 2				
	<u>Spec</u>			<u>Pass/Fail</u>	<u>Spec</u>			<u>Pass/Fail</u>	
HDL mg/dL	32	-	54	Pass	HDL mg/dL	70	-	104	Pass
TCHOL mg/dL	145	-	215	Pass	TCHOL mg/dL	267	-	383	Pass
TGL mg/dL	56	-	132	Pass	TGL mg/dL	314	-	454	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

1/5/2026  
Date