



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Lactate Sensor
Lot Number 26007011

Part Number 48553
Exp Date 2028-01-07
Date of Manufacture 2026-01-07
UDI# (01)00385480485531(11)260107(17)280107(10)26007011

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	77 - 91	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

1/20/2026
Date