



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose and β -Ketone Control Solution Level 1

Nova Reference 46947
LifeScan Reference 02312801
Lot Number 0425330301

Exp Date 2028-05-26
Date of Manufacture 2025-11-26
UDI# (01)00385480469470(11)251126(17)280526(10)0425330301

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	62
Ketone mmol/L	0.1 - 0.2	0.1

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

1/16/2026
Date