



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Solution Level 2
LifeScanReference 02307801
Lot Number 0425017302

Ref Number 41742
Date of Manufacture 2025-01-17
Exp Date 2027-07-17
UDI# (01)00385480417426(11)250117(17)270717(10)0425017302

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	105 - 125	115

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector
Title

1/27/2026
Date