



**CERTIFICATE OF ANALYSIS**

STP Prime Plus Vet Reference Cartridge  
Lot Number 25358014

Ref Number 59345  
Exp Date 2027-06-24  
Date of Manufacture 2025-12-24

Internal QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	136.7 - 144.7	Pass
K mmol/L	3.69 - 4.19	Pass
CL mmol/L	122.0 - 131.0	Pass

External QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	134.3 - 142.3	Pass
K mmol/L	3.73 - 4.23	Pass
CL mmol/L	118.9 - 127.9	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

1/27/2026  
Date