



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 Chemistry Calibrator
and Reagent Cartridge MV
Lot Number 25364100

Ref Number 58101
Exp Date 2026-12-12
Date of Manufacture 2025-12-12

Chemistry Cal Cartridge L/N: 25364097 Exp Date: 2027-03-17

Cal D1 25351007

Cal D2 25353010

Table with 3 columns: Analyte, Spec, Result. Rows include Gln mmol/L, Glu mmol/L, Gluc g/L, Lac g/L, NH4 mmol/L, Na+ mmol/L, K+ mmol/L, Ca++ mmol/L.

Table with 3 columns: Analyte, Spec, Result. Rows include Gln mmol/L, Glu mmol/L, Gluc g/L, Lac g/L, NH4 mmol/L, Na+ mmol/L, K+ mmol/L, Ca++ mmol/L.

Chemistry Reagent Cartridge L/N: 25364093 Exp Date: 2026-12-12

PF Soln L/N 25349048

R Soln L/N 25350043

Table with 3 columns: Analyte, Spec, Result. Row includes pH.

Table with 3 columns: Analyte, Spec, Pass/Fail. Row includes KCl mol/L.

DP Soln L/N 25346055

DF Soln L/N 25350041

Table with 3 columns: Analyte, Spec, Pass/Fail. Row includes NaOCl %.

Table with 3 columns: Analyte, Spec, Pass/Fail. Row includes pH.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi-M

Approval

QC Inspector
Title

1/28/2026
Date