CERTIFICATE OF ANALYSIS

STP pHOx Ultra/ CCX BUN Sensor
Lot Number 25365040

Part Number 48555
Exp Date 2027-03-31
Date of Manufacture 2025-12-31
UDI# (01)00385480485555(11)251231(17)270331(10)25365040

	Spec	Pass/Fail
Slope	≥ 9.60	Pass
pHOx Ctl L4	14-24	Pass
pHOx Ctl L5	49-63	Pass
Std D mV	0.00 ± 60.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

1/20/2026
Date