



CERTIFICATE OF ANALYSIS

STP Prime Plus Vet Reference Cartridge
Lot Number 26016078

Ref Number 59345
Exp Date 2027-07-16
Date of Manufacture 2026-01-16

| Internal QC Level 4 | <u>Spec</u> | <u>Pass/Fail</u> |
|---------------------|---------------|------------------|
| Na mmol/L | 136.7 - 144.7 | Pass |
| K mmol/L | 3.69 - 4.19 | Pass |
| CL mmol/L | 122.0 - 131.0 | Pass |

| External QC Level 4 | <u>Spec</u> | <u>Pass/Fail</u> |
|---------------------|---------------|------------------|
| Na mmol/L | 134.3 - 142.3 | Pass |
| K mmol/L | 3.73 - 4.23 | Pass |
| CL mmol/L | 118.9 - 127.9 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

1/28/2026
Date