



**CERTIFICATE OF ANALYSIS**

BioProfile FLEX 2 CDV Reagent Pack  
Lot Number 25356041

Ref Number 55297  
Exp Date 2026-11-24  
Date of Manufacture 2025-11-24

	CDV Bottle Pack L/N:	25356040	Exp Date:	2026-11-24		
Soln SF L/N	25329026		Soln DP L/N	25336052		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
pH	6.900 - 7.100	Pass	Depro Soln	N/A	Present	
Cal C L/N	25328030		Soln T L/N	25328027		
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>	
3% Detergent	N/A	Present	0.4% Trypan	N/A	Present	

CDV Reagent Cartridge L/N: 25356038      Exp Date: 2027-03-01  
 Soln PF L/N      25335025  
                          Spec      Pass/Fail  
 pH      6.900 - 7.100      Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Awishi-M*

Approval

QC Inspector  
Title

1/30/2026  
Date