



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 400 Sample
Lot Number 25364079

Ref Number 53463
Exp Date 2027-04-17
Date of Manufacture 2025-10-17
UDI# (01)00385480534635(11)251017(17)270417(10)25364079

Cal A	L/N	25290020	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.354	N/A
	pCO2 mmHg	24.7	N/A

Cal B	L/N	25290019	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	25346041	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	53.1	N/A
	pO2 mmHg	103.0	N/A

Reference	L/N	25290018	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishim

Approval

QC Inspector
Title

01/28/2026
Date