



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 500 Sample
Lot Number 25357053

Ref Number 53464
Exp Date 2027-03-16
Date of Manufacture 2025-09-16
UDI# (01)00385480534642(11)250916(17)270316(10)25357053

Cal A L/N 25259045		<u>Assigned Value*</u>	<u>Result</u>
pH		7.352	N/A
pCO2 mmHg		24.9	N/A

Cal B L/N 25259046		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush L/N 25346041		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		53.1	N/A
pO2 mmHg		103.0	N/A

Reference L/N 25259047		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector
Title

01/27/2026
Date