



CERTIFICATE OF ANALYSIS

STP Prime Plus Calibrator Cartridge 500 Sample
Lot Number 25358026

Ref Number 57829
Exp Date 2027-06-15

Date of Manufacture 2025-12-15

UDI# (01)00385480578295(11)251215(17)270615(10)25358026

Table with 3 columns: Cal A L/N 25349044, Cal B L/N 25349045, Cal C L/N 25349046. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, tHb g/dL, and Assigned Value*/Result.

Table with 3 columns: Cal D L/N 25349047, Flush L/N 25350025, Reference L/N 25349043. Rows include pH @ 37C, PO2, and KCl.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector
Title

1/30/2026
Date