



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 400 Sample w. Creat/BUN
Lot Number 25350031

Ref Number 57834
Exp Date 2027-05-21

Date of Manufacture 2025-11-21

UDI# (01)00385480578349(11)251121(17)270521(10)25350031

Table with 3 columns: Cal A L/N 25339029, Cal B L/N 25339030, Cal C L/N 25339031. Rows include HCO3, Na, K, Cl, Ca, Mg, Glu, Lac, BUN, Creat with Spec and Pass/Fail columns.

Table with 3 columns: Cal D L/N 25339032, Flush L/N 25325033, Reference L/N 25339026. Rows include pH @ 37C, PO2, KCl with Spec and Pass/Fail columns.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector
Title

1/30/2026
Date