



**CERTIFICATE OF ANALYSIS**

Bioprofile Flex Linearity Solution LOW  
Lot Number 25344020

Catalog Number 45278  
Exp Date 2027-06

|            | Spec      | Pass/Fail |
|------------|-----------|-----------|
| Gln mmol/L | 0.35-0.65 | Pass      |
| Glu mmol/L | 0.35-0.65 | Pass      |
| Gluc g/L   | 0.35-0.65 | Pass      |
| Lac g/L    | 0.35-0.65 | Pass      |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Gluc, Lac

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "S. D. Long", is written over a light blue horizontal line.

Approval

QC Inspector  
Title

1/19/2026  
Date