



## CERTIFICATE OF ANALYSIS

Bioprofile pH Linearity Solution HIGH  
Lot Number 25349011

Catalog Number 49938  
Exp Date 2026-12

	<u>Spec</u>	<u>Pass/Fail</u>
pH	7.725 - 7.925	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacture

*Awishi.M*

Approval

QC Inspector  
Title

1/22/2026  
Date