



## CERTIFICATE OF ANALYSIS

Nova 5 Calibrator Pack  
Lot Number 531118

Ref Number 15648  
Exp Date 2027-11-07  
Date of Manufacture 2025-11-07  
UDI# (01)00385480156486(11)251107(17)271107(10)531118

Cal A L/N	25311018	Cal B L/N	25311018	Reference L/N	25311018
Na mmol/L	139.0 - 141.0	Spec	Pass/fail	Na mmol/L	59.4 - 60.6
K mmol/L	3.97 - 4.03	Pass		K mmol/L	9.8 - 10.2
Cl mmol/L	113.6 - 115.4	Pass		Cl mmol/L	45.5 - 46.5
				KCl mol/L	1.8 - 2.2
					Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

*Awishi.M*

Approval

QC Inspector  
Title

1/22/2026  
Date