



CERTIFICATE OF ANALYSIS

Allegro Control 1 UACR
Lot Number 8526027241

Ref Number 51032
Exp Date 2028-01-27
Date of Manufacture 2026-01-27
UDI# (01)00385480510325(11)260127(17)280127(10)8526027241

	Spec	Pass/Fail
Albumin, mg/L	33-81	Pass
Creatinine, mg/dL	65-119	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "S. Smith", is located above the approval section.

Approval

QC Inspector
Title

2/4/2026
Date