



**CERTIFICATE OF ANALYSIS**

Allegro HbA1c Test Cartridge  
Lot Number 25356009

Ref Number 54215  
Exp Date 2026-12-22  
Date of Manufacture 2025-12-22  
UDI# (01)00385480542159(11)251222(17)261222(10)25356009

Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
HbA1c %	4.8 - 6.6	Pass	HbA1c %	6.8 - 10.4	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

1/7/2026  
Date