



CERTIFICATE OF ANALYSIS

CDV Analyzer Reagent Pack V2
Lot Number 26034003

Ref Number 58117
Exp Date 2027-01

	CDV Bottle Pack L/N: 26034002		Exp Date: 2027-01		
C Soln L/N	26014051		T Soln L/N	26023010	
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
3% Detergent	N/A	Present	0.4% Trypan	N/A	Present
pH	> 9.0	Pass	Particle Count	≤ 2.0	Pass

	CDV Cartridge L/N: 26034001		Exp Date: 2027-01		
S Soln L/N	26014050		D Soln L/N	26014049	
	<u>Spec</u>	<u>Result</u>		<u>Spec</u>	<u>Result</u>
pH	6.900 - 7.100	Pass	%NaOCl	0.43 - 0.73	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/9/2026
Date