



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, NOVA MAX PRO CREATININE eGFR, 50 COUNT
Catalog Number:	63983
Nova Biomedical P/N	63983
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10%	PASS FAIL PASS FAIL
Qty:	1,620
Lot No.	2730226023
Mfg. Date:	2026-01-23
Expiration Date:	2027-01-23
Unique Device Identifier (UDI):	(01)20385480639839 (11)260123 (17)270123 (10)2730226023

Mark Bowene
 Quality Control Manager

2/3/2026
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454