



**CERTIFICATE OF ANALYSIS**

STP Prime ABG Cal Cartridge 500 Sample  
Lot Number 26014032

Ref Number 53464  
Exp Date 2027-06-30  
Date of Manufacture 2025-12-31  
UDI# (01)00385480534642(11)251231(17)270630(10)26014032

Cal A	L/N	26002022	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.351	N/A
pCO2 mmHg		25.3	N/A

Cal B	L/N	26006018	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	25365023	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		53.3	N/A
pO2 mmHg		100.5	N/A

Reference	L/N	26002024	
		<u>Spec</u>	<u>Pass/Fail</u>
KCI		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

02/12/2026  
Date