



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0426027301

Exp Date 2028-07-27

Date of Manufacture 2026-01-27

UDI # (01)00385480417419(11)260127(17)280727(10)0426027301

	Spec	Mean Result
Glucose mg/dL	57 - 65	59

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "M. Knight".

Approval

QC Inspector

Title

2/10/2026

Date