



CERTIFICATE OF ANALYSIS

Linearity Standard Set G Multipack
Lot Number 26008042

Ref Number 56198
Exp Date 2028-01-06
Date of Manufacture 2026-01-06
UDI# (01)00385480561983(11)260106(17)280106(10)26008042

Level 1 L/N 26006025
Level 1 Exp Date 2028-01-06

	<u>Spec</u>	<u>Pass/Fail</u>
Hct %	18 - 22	Pass
Hb g/dL	5.3 - 7.3	Pass

Level 2 L/N 26006024
Level 1 Exp Date 2028-01-06

	<u>Spec</u>	<u>Pass/Fail</u>
Hct %	56 - 60	Pass
Hb g/dL	17.5 - 20.5	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

2/9/2026

Date