nova biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	2,520
Lot No.	4926040129
Mfg. Date:	2026-02-09
Expiration Date:	2027-02-09
Unique Device Identifier (UDI):	(01)20385480432720 (11)260209 (17)270209 (10)4926040129

Marc Browne 2/11/26

Quality Control Manager Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454

Released Confidential