



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Level 3

Nova Reference 41743
LifeScan Reference 02315601
Lot Number 0426021303

Exp Date 2028-07-21
Date of Manufacture 2026-01-21
UDI#(01)00385480417433(11)260121(17)280721(10)0426021303

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	304

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/13/2026
Date