



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Hb/Hct Control Level 1  
Lot Number 2026035241

Ref Number 59820  
Exp Date 2028-02-04  
Date of Manufacture 2026-02-04  
UDI#(01)00385480598200(11)260204(17)280204(10)2026035241

	<u>Spec</u>	<u>Mean Result</u>
Hb g/dL	8.2 - 8.8	8.7
Hct %	24 - 26	26

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

2/12/2026  
Date