



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 100 Sample
Lot Number 26034042

Ref Number 53104
Exp Date 2027-07-14
Date of Manufacture 2026-01-14
UDI# (01)00385480531047(11)260114(17)270714(10)26034042

Cal A	L/N	26016056	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.347	N/A
	pCO2 mmHg	25.0	N/A

Cal B	L/N	26016057	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A

Flush	L/N	26014014	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	52.4	N/A
	pO2 mmHg	102.8	N/A

Reference	L/N	26016058	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCI	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

02/12/2026
Date