



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 100 Sample
Lot Number 25357053

Ref Number 53104
Exp Date 2027-03-16
Date of Manufacture 2025-09-16
UDI# (01)00385480531047(11)250916(17)270316(10)25357053

Cal A	L/N	25259045		
		<u>Assigned Value*</u>	<u>Result</u>	
	pH	7.352	N/A	
	pCO2 mmHg	24.9	N/A	

Cal B	L/N	25259046		
		<u>Spec</u>	<u>Pass/Fail</u>	
	pH	6.837 - 6.843	Pass	
	pO2 mmHg	167.0	N/A	

Flush	L/N	25346041		
		<u>Assigned Value*</u>	<u>Result</u>	
	pCO2 mmHg	53.1	N/A	
	pO2 mmHg	103.0	N/A	

Reference	L/N	25259047		
		<u>Spec</u>	<u>Pass/Fail</u>	
	KCI	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

02/12/2026

Date