



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Linearity Kit
Lot Number 0526009990

Ref Number 42173

Exp Date 2028-07-09

Date of Manufacture 2026-01-09

UDI# (01)00385480421737(11)260109(17)280709(10)0526009990

Level 1 L/N	0526013301
Level 1 Exp Date	2028-07-13
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	16 - 22 18

Level 2 L/N	0526009302
Level 2 Exp Date	2028-07-09
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	57 - 65 60

Level 3 L/N	0526014303
Level 3 Exp Date	2028-07-14
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	105 - 125 112

Level 4 L/N	0526013304
Level 4 Exp Date	2028-07-13
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	275 - 325 301

Level 5 L/N	0526009305
Level 5 Exp Date	2028-07-09
	<u>Spec</u> <u>Mean Result</u>
Glucose mg/dL	470 - 540 512

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. Analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/13/2026
Date