



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Ultra/CCX Na Sensor
Lot Number 26019031

Ref Number 48579
Exp Date 2027-07-19

Date of Manufacture 2026-01-19

UDI# (01)00385480485791(11)260119(17)270719(10)26019031

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.30 - 11.00	Pass
Std C mV	-10.00 - +70.00	Pass
L4 CtL	136.0 - 144.0	Pass
L5 CtL	110.5 - 118.5	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/11/2026
Date