



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 200 Sample
Lot Number 26022042

Ref Number 53360
Exp Date 2027-07-09
Date of Manufacture 2026-01-09
UDI# (01)00385480533607(11)260109(17)270709(10)26022042

Cal A	L/N	26014011		
		<u>Assigned Value*</u>	<u>Result</u>	
	pH	7.346	N/A	
	pCO2 mmHg	25.4	N/A	

Cal B	L/N	26014012		
		<u>Spec</u>	<u>Pass/Fail</u>	
	pH	6.837 - 6.843	Pass	
	pO2 mmHg	167.0	N/A	

Flush	L/N	26009056		
		<u>Assigned Value*</u>	<u>Result</u>	
	pCO2 mmHg	52.2	N/A	
	pO2 mmHg	102.3	N/A	

Reference	L/N	26014013		
		<u>Spec</u>	<u>Pass/Fail</u>	
	KCI	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector

Title

02/26/2026

Date