



CERTIFICATE OF ANALYSIS

STP pH_{OX}/pH_{OX} Basic/BiopH_{OX} Air Detector
Lot Number 26033023

Ref Number 21513
Date of Manufacture 2026-02-02

	<u>Spec</u>	<u>Pass/Fail</u>
AD1 mV	1400 - 2600	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in blue ink that reads "Orlin Rammer".

Approval

QC Inspector

Title

2/11/2026

Date