



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 pH/Gas Reference Sensor
Lot Number 26034038

Part Number 57467
Exp Date 2027-08-03
Date of Manufacture 2026-02-03

| | <u>Spec*</u> | <u>Pass/Fail</u> |
|----------------|---------------|------------------|
| pH Slope | 8.70 - 11.80 | Pass |
| BioPro Ctl L 1 | 6.628 - 6.728 | Pass |
| BioPro Ctl L 3 | 7.328 - 7.428 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector
Title

02/12/2026
Date