



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge 100 Sample
 Lot Number 26041083

Ref Number 52861
 Exp Date 2027-07-16
 Date of Manufacture 2026-01-16
 UDI# (01)00385480528610(11)260116(17)270716(10)26041083

Cal A	L/N	26028086	
		<u>Assigned Value*</u>	<u>Result</u>
	pH	7.356	N/A
	pCO2 mmHg	25.0	N/A
		<u>Spec</u>	<u>Pass/Fail</u>
	Na mmol/L	132.1 - 133.9	Pass
	K mmol/L	3.97 - 4.03	Pass
	Cl mmol/L	105.2 - 106.8	Pass
	iCa mmol/L	1.08 - 1.12	Pass
	Glu mg/dL	79 - 81	Pass
	Lac mmol/L	1.8 - 2.2	Pass

Cal B	L/N	26028087	
		<u>Spec</u>	<u>Pass/Fail</u>
	pH	6.837 - 6.843	Pass
	pO2 mmHg	167.0	N/A
	Na mmol/L	68.0 - 69.0	Pass
	K mmol/L	9.92 - 10.08	Pass
	Cl mmol/L	45.6 - 46.4	Pass
	iCa mmol/L	2.17 - 2.23	Pass
	Glu mg/dL	197 - 203	Pass
	Lac mmol/L	9.5 - 10.5	Pass

Flush	L/N	26016054	
		<u>Assigned Value*</u>	<u>Result</u>
	pCO2 mmHg	53.0	N/A
	pO2 mmHg	104.6	N/A

Reference	L/N	26023012	
		<u>Spec</u>	<u>Pass/Fail</u>
	KCl	1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi-M

Approval

QC Inspector
 Title

02/26/2026
 Date