



**CERTIFICATE OF ANALYSIS**

Prime Reference Cartridge  
Lot Number 26023033

Part Number 42043  
Exp Date 2027-07-23  
Date of Manufacture 2026-01-23  
UDI# (01)00385480420433(11)260123(17)270723(10)26023033

|           | <u>Spec</u>  | <u>Pass/Fail</u> |
|-----------|--------------|------------------|
| Na+ Slope | 9.50 - 12.00 | Pass             |
| K+ Slope  | 9.00 - 11.50 | Pass             |
| Hct Slope | 2.00 - 3.00  | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Scott Quinn".

Approval

QC Inspector  
Title

2/6/2026  
Date