



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 Chemistry Calibrator
and Reagent Cartridge LV
Lot Number 26055083

Ref Number 58003
Exp Date 2027-01-06
Date of Manufacture 2026-01-06

Chemistry Cal Cartridge L/N: 26055080 Exp Date: 2027-04-07

Cal D1 26007012

Cal D2 26009012

Table with 3 columns: Analyte, Spec, Result. Rows include Gln mmol/L, Glu mmol/L, Gluc g/L, Lac g/L, NH4 mmol/L, Na+ mmol/L, K+ mmol/L, Ca++ mmol/L.

Table with 3 columns: Analyte, Spec, Result. Rows include Gln mmol/L, Glu mmol/L, Gluc g/L, Lac g/L, NH4 mmol/L, Na+ mmol/L, K+ mmol/L, Ca++ mmol/L.

Chemistry Reagent Cartridge L/N: 26055076 Exp Date: 2027-01-06

PF Soln L/N 26021050

R Soln L/N 26006061

Table with 3 columns: Analyte, Spec, Result. Row: pH 6.85 - 7.05 Pass

Table with 3 columns: Analyte, Spec, Pass/Fail. Row: KCl mol/L 1.8 - 2.2 Pass

DP Soln L/N 26006034

DF Soln L/N 2013013

Table with 3 columns: Analyte, Spec, Pass/Fail. Row: NaOCl % 0.43 - 0.73 Pass

Table with 3 columns: Analyte, Spec, Pass/Fail. Row: pH 6.90 - 7.10 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Handwritten signature

Approval

QC Inspector

Title

3/4/2026

Date