



## CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX pH Sensor  
Lot Number 26014027

Part Number 48580  
Exp Date 2027-07-14  
Date of Manufacture 2026-01-14  
UDI# (01)00385480485807(11)260114(17)270714(10)26014027

|          | <u>Spec</u>       | <u>Pass/Fail</u> |
|----------|-------------------|------------------|
| Slope    | 9.20 - 11.20      | Pass             |
| Std C mV | 0.00 $\pm$ 110.00 | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith" or similar, written in a cursive style.

Approval

QC Inspector  
Title

2/12/2026  
Date