




CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A		
Description:	TEST STRIPS, STATSTRIP CAPILLARY LACTATE 50 COUNT, PACKAGED		
Catalog Number:	61440		
Nova Biomedical P/N	61440		
<u>Target Range</u>	<u>CV (%)</u>	<u>SD (mmol/L)</u>	
0.8-2.5 mmol/L	≤ 6.5%	±0.3	PASS / FAIL
3.0-5.0 mmol/L	≤ 5.5%	±0.5	PASS / FAIL
10-14 mmol/L	≤ 5.0%	±1.2	PASS / FAIL
Qty:	1,497		
Lot No.	2626055249		
Mfg. Date:	2026-02-24		
Expiration Date:	2028-02-24		
Unique Device Identifier (UDI):	(01)20385480614409 (11) 260224 (17) 280224 (10) 2626055249		


 Quality Control Manager

3/4/26
 Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454