



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX pH Sensor
Lot Number 26044033

Part Number 48580
Exp Date 2027-08-13
Date of Manufacture 2026-02-13
UDI# (01)00385480485807(11)260213(17)270813(10)26044033

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.20 - 11.20	Pass
Std C mV	0.00 ± 110.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "Romanus Achten".

Approval

QC Inspector
Title

3/9/2026
Date