



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Ionized Mg Sensor
Lot Number 26050031

Part Number 48556
Exp Date 2027-05-19
Date of Manufacture 2026-02-19
UDI# (01)10385480485569(11)260219(17)270519(10)26050031

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------------|------------------|------------------|
| Slope | 9.70 - 14.00 | Pass |
| Std C mV | -100.00 - +60.00 | Pass |
| pHOx Ctl L4 | 0.55 - 0.73 | Pass |
| pHOx Ctl L5 | 1.00 - 1.30 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

3/10/2026
Date