



**CERTIFICATE OF ANALYSIS**

StatPro pHox/CCX SO2 Calibrator

Ref Number 22771

Lot Number 26043017

Exp Date 2027-07-22

Date of Manufacture 2026-01-22

UDI# (01)00385480227711(11)260122(17)270722(10)26043017

Level 1 L/N: 26022010

Level 2 L/N: 26022011

Level 1 Exp Date: 2027-07-22

Level 2 Exp Date: 2027-07-22

	<u>Spec</u>	<u>Pass/Fail</u>
SO2 %	45 - 65	Pass
Hct %	51 - 55	Pass

	<u>Spec</u>	<u>Pass/Fail</u>
SO2 %	92 - 105	Pass
Hct %	51 - 55	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials. SO2 is traced to tonometry.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi M*

Approval

QC Inspector

Title

3/13/2026

Date