



**CERTIFICATE OF ANALYSIS**

Bioprofile pH Linearity Solution LOW  
Lot Number 26021014

Catalog Number 49939  
Exp Date 2027-01

	Spec	Pass/Fail
pH	6.200 - 6.400	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in cursive script, appearing to read "Sandang".

Approval

QC Inspector  
Title

2/10/2026  
Date