



CERTIFICATE OF ANALYSIS

STP pHox/Basic Controls Auto-Cartridge
Lot Number 26051044

Ref Number 23931
Exp Date 2027-08-06
Date of Manufacture 2026-02-06
UDI#(01)00385480239318(11)260206(17)270806(10)26051044

Table with 3 columns: Parameter, Spec, Pass/Fail. Rows include Level 1 (pH, pCO2, pO2, SO2, Hct), Level 2 (pH, pCO2, pO2), and Level 3 (pH, pCO2, pO2, SO2, Hct).

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials. SO2 is traced to tonometry.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer"

Awishi.M

Approval

QC Inspector

Title

3/10/2026

Date