



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Ultra/CCX pO2 Sensor
Lot Number 26041043

Ref Number 48598

Exp Date 2028-02-10

Date of Manufacture 2026-02-10

UDI# (01)00385480485982(11)260210(17)280210(10)26041043

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector
Title

3/6/2026
Date