



**CERTIFICATE OF ANALYSIS**

Stat Profile pHox Ultra / CCX SO2 Sensor  
Lot Number 26063025

Part Number 48591  
Exp Date 2029-03-04  
Date of manufacture 2026-03-04  
UDI#(01)00385480485913(11)260304(17)290304(10)26063025

|       | <u>Spec</u>  | <u>Pass/Fail</u> |
|-------|--------------|------------------|
| Slope | 7.00 - 14.00 | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector  
Title

3/10/2026  
Date