



CERTIFICATE OF ANALYSIS

Preheater Cleaning Agent
Lot Number 26055015

Ref Number 11802
Exp Date 2027-02-24
Date of Manufacture 2026-02-24
UDI# (01)00385480118026(11)260224(17)270224(10)26055015

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.43 - 0.73%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in dark ink, appearing to read "M. Knight", is written over a horizontal line.

Approval

QC Inspector
Title

3/12/2026
Date